

Product Description: Accu-Chek Comfort Curve Blood Glucose Test Strips
Accu-Chek Advantage Blood Glucose Test Strips

Manufacturer: Roche Diagnostics

UPCs: 0-75537-00553-8
0-75537-30373-3
0-75537-30381-8

Date Recall Posted: 07/19/06

Summary Of Recall: Roche Diagnostics has become aware that in some vials of the affected test strips, the drying agent beads that normally are located in the cap of the vial have become loose. If loose drying agent beads are discarded the test strips are no longer protected against moisture, potentially leading to an inaccurate result.

What You Should Do: If a customer notices loose drying agent in their vial, they are to contact Roche Diagnostics directly at the number provided:

Accu-Chek Customer Care Service Center 1-800-525-8072